

# MET LABORATORIES, INC. CLASSIFICATION RECORD



*The applicant named below has been authorized by MET Laboratories, Inc. to represent the product(s) listed in this record as "MET Classified" and to mark this/these product(s) according to the terms and conditions of the MET Mark Utilization Agreement, MET Listing Reports, and the applicable marking agreements. Only the product(s) bearing the MET Mark and under a follow-up service are considered to be included in the MET Classification program.*

**FILE NUMBER:**

E212246

**APPROVAL DATE:**

April 25, 2005

**REVISED:**

-

PRODUCT(S)	MODEL(S)	ELECTRICAL RATINGS
Infrared Energy Therapy System	480	100-240V, 1.6A, 47-63Hz

STANDARD NUMBER	STANDARD TITLE	EDITION
UL60601-1	Medical Equipment	1st
CSA 22.2 No 601-1	Medical Equipment	M90-2001

MET LABORATORIES, INC. requires that any and all changes proposed in the previously identified product(s), that affects the information contained in the above referenced listing report, must be submitted to MET for evaluation prior to implementation to assure continued MET Classification status.

The above identified product(s) has/have been submitted by the applicant:

**APPLICANT:**

Anodyne Therapy Systems, LLC  
13570 Right Circle  
Tampa, FL 33626

The covered products shall be subjected to quarterly follow-up inspections to ensure that the Classified product(s) are identical to the representative product sample evaluated by MET LABORATORIES, INC. and that all manufacturer's responsibilities are being fulfilled as specified in the MANUFACTURING RESPONSIBILITY section of the Certification report.

**CLASSIFICATION CONDITIONS: \***

**\* The products were not evaluated to comply with clause 52.1 (Programmable Electronic Systems) and clause 48 (Biocompatibility).**

Roger Smith  
Sr. Project Engineer  
Safety Laboratory

Rick Cooper  
Director of Laboratory Operations,  
Safety Laboratory



*MET Laboratories, Inc. is accredited by OSHA and the Standards Council of Canada.  
The Nation's First Nationally Recognized Testing Laboratory*

**NRTL**